

MEDICAL ELECTRONIC SYSTEMS

WWW.MES-GLOBAL.COM

QwikCheck™ Beads Batch Release Form									
BATCH #		Level 1			Level 2			NEGATIVE CONTROL	
EXP DATE:	15-Apr-13	LOT#	150412001	LOT#	150412002	LOT#	150412003		
For use with:		Conc.			Conc.			Conc	MSC
SQA-V Automated		44	+/-	6.2	22	+/-	4.4	0.0	0.0
SQA-V Visualization		45	+/-	11.3	23	+/-	5.8	0.0	0.0
Hemocytometer		45	+/-	11.3	23	+/-	5.8	0.0	NA
Makler		44	+/-	11	22	+/-	5.5	0.0	NA
Fixed Coverslip		44	+/-	11	23	+/-	5.8	0.0	NA