

**MEDICAL ELECTRONIC SYSTEMS**

**WWW.MES-GLOBAL.COM**

QwikCheck™ Beads Batch Release Form									
BATCH #	150112	Level 1			Level 2			NEGATIVE CONTROL	
EXP DATE:	15-01-2013	LOT#	150112001	LOT#	150112002	LOT#	150112003		
For use with:		Conc.			Conc.			Conc	MSC
SQA-V Automated		45	+/-	6.3	22	+/-	4.4	0.0	0.0
SQA-V Visualization		47	+/-	11.8	23	+/-	5.8	0.0	0.0
Hemocytometer		46	+/-	11.5	23	+/-	5.8	0.0	NA
Makler		46	+/-	11.5	24	+/-	6.0	0.0	NA
Fixed Coverslip		45	+/-	11.3	23	+/-	5.8	0.0	NA