

MEDICAL ELECTRONIC SYSTEMS

WWW.MES-GLOBAL.COM

QwikCheck™ Beads Batch Release Form									
BATCH #	090713	Level 1			Level 2			NEGATIVE CONTROL	
EXP DATE:	7/9/2014	LOT#	090713001	LOT#	090713002	LOT#	090713003		
For use with:	Conc.			Conc.			Conc	MSC	
SQA-V Automated	44	+/-	6.2	22	+/-	4.4	0.0	0.0	
SQA-V Visualization	45	+/-	11.3	22	+/-	5.5	0.0	0.0	
Hemocytometer	44	+/-	11.0	23	+/-	5.8	0.0	NA	
Makler	47	+/-	11.8	23	+/-	5.8	0.0	NA	
Fixed Coverslip	45	+/-	11.3	23	+/-	5.8	0.0	NA	