

MEDICAL ELECTRONIC SYSTEMS
WWW.MES-GLOBAL.COM

QwikCheck™ Beads Batch Release Form									
BATCH #	011113	Level 1			Level 2			NEGATIVE CONTROL	
EXP DATE:	01/11/2014	LOT#	011113001	LOT#	011113002	LOT#	011113003		
For use with:	Conc.			Conc.			Conc.	MSC	
SQA-V Automated	45	+/-	6.3	23	+/-	4.6	0.0	0.0	
SQA-V Visualization	44	+/-	11.0	22	+/-	5.5	0.0	0.0	
Hemocytometer	44	+/-	11.0	22	+/-	5.5	0.0	NA	
Makler	45	+/-	11.3	22	+/-	5.5	0.0	NA	
Fixed Coverslip	45	+/-	11.3	23	+/-	5.8	0.0	NA	

Marina Friling

[Signature]
Initials: Level 1

[Signature]
Initials: Level 2

[Signature]
Initials: NEG

Dr. Lev Rabinovitch, CTO

[Signature]
Initials: Level 1

[Signature]
Initials: Level 2

[Signature]
Initials: NEG

Marcia Deutsch, CEO

[Signature]
Initials: Level 1

[Signature]
Initials: Level 2

[Signature]
Initials: NEG

3/11/13
Date

3/11/13
Date

3/11/13
Date