

MEDICAL ELECTRONIC SYSTEMS

WWW.MES-GLOBAL.COM

QwikCheck™ Beads Batch Release Form									
BATCH #	280414	Level 1			Level 2			NEGATIVE CONTROL	
EXP DATE:	28_April_15	LOT# 280414001			LOT# 280414002			LOT# 280414003	
For use with:		Conc.			Conc.			Conc	MSC
SQA-V Automated		44	+/-	6.2	22	+/-	4.4	0.0	0.0
SQA-V Visualization		44	+/-	11	21	+/-	5.3	0.0	0.0
Hemocytometer		45	+/-	11.3	22	+/-	5.5	0.0	NA
Makler		46	+/-	11.5	23	+/-	5.8	0.0	NA
Fixed Coverslip		44	+/-	11	22	+/-	5.5	0.0	NA

* Makler chamber was loaded with 10ul.