

MEDICAL ELECTRONIC SYSTEMS  
WWW.MES-GLOBAL.COM

QwikCheck™ Beads Batch Release Form									
BATCH #	011214	Level 1			Level 2			NEGATIVE CONTROL	
EXP DATE:	01_December_15	LOT# 010914001			LOT# 011214002			LOT# 011214003	
For use with:	Conc.			Conc.			Conc	MSC	
SQA-V Automated	44	+/-	6.2	22	+/-	4.4	0.0	0.0	
SQA-V Visualization	44	+/-	11.0	22	+/-	5.5	0.0	0.0	
Hemocytometer	44	+/-	11.0	23	+/-	5.8	0.0	NA	
Makler	45	+/-	11.3	23	+/-	5.8	0.0	NA	
Fixed Coverslip	43	+/-	10.8	22	+/-	5.5	0.0	NA	

Makler chamber was loaded with 5ul