

MEDICAL ELECTRONIC SYSTEMS

WWW.MES-GLOBAL.COM

QwikCheck™ Beads Batch Release Form									
BATCH #	070715	Level 1			Level 2			NEGATIVE CONTROL	
EXP DATE:	7-Jul-16	LOT# 070715001			LOT# 070715002			LOT# 070715003	
For use with:	Conc.			Conc.			Conc	MSC	
SQA / QwikCheck gold Automated	46	+/-	6.4	24	+/-	4.8	0.0	0.0	
SQA-V Visualization	46	+/-	11.5	24	+/-	6.0	0.0	0.0	
Hemocytometer	46	+/-	11.5	24	+/-	6.0	0.0	NA	
Makler	46	+/-	11.5	24	+/-	6.0	0.0	NA	
Fixed Coverslip	46	+/-	11.5	24	+/-	6.0	0.0	NA	