

MEDICAL ELECTRONIC SYSTEMS

WWW.MES-GLOBAL.COM

QwikCheck™ Beads Batch Release Form									
BATCH #	221215	Level 1			Level 2			NEGATIVE CONTROL	
EXP DATE:	22-Dec-16	LOT# 221215001			LOT# 221215002			LOT# 221215003	
For use with:	Conc.			Conc.			Conc	MSC	
SQA / QwikCheck gold Automated	47	+/-	6.6	26	+/-	5.2	0.0	0.0	
SQA-V Visualization	46	+/-	11.5	26	+/-	6.5	0.0	NA	
Hemocytometer	46	+/-	11.5	26	+/-	6.5	0.0	NA	
Makler	46	+/-	11.5	26	+/-	6.5	0.0	NA	
Fixed Coverslip	48	+/-	12.0	26	+/-	6.5	0.0	NA	