

**MEDICAL ELECTRONIC SYSTEMS**

**WWW.MES-GLOBAL.COM**

QwikCheck™ Beads Batch Release Form									
BATCH #	120416	Level 1			Level 2			NEGATIVE CONTROL	
EXP DATE:	12-Apr-17	LOT#	120416001	LOT#	120416002	LOT#	120416003		
For use with:	Conc.			Conc.			Conc	MSC	
SQA / QwikCheck gold Automated	48	+/-	6.7	26	+/-	5.2	0.0	0.0	
SQA-V Visualization	49	+/-	12.3	24	+/-	6.0	0.0	NA	
Hemocytometer	48	+/-	12.0	25	+/-	6.3	0.0	NA	
Makler	49	+/-	12.3	27	+/-	6.8	0.0	NA	
Fixed Coverslip	49	+/-	12.3	26	+/-	6.5	0.0	NA	