

**MEDICAL ELECTRONIC SYSTEMS**

**WWW.MES-GLOBAL.COM**

<b>QwikCheck™ Beads Batch Release Form</b>									
<b>BATCH #</b>	050716	<b>Level 1</b>			<b>Level 2</b>			<b>NEGATIVE CONTROL</b>	
<b>EXP DATE:</b>	5-Jul-17	LOT# 050716001			LOT# 050716002			LOT# 050716003	
<b>For use with:</b>		<b>Conc.</b>			<b>Conc.</b>			<b>Conc</b>	<b>MSC</b>
<b>SQA / QwikCheck gold Automated</b>		49	+/-	6.9	24	+/-	4.8	0.0	0.0
<b>SQA-V Visualization</b>		50	+/-	12.5	24	+/-	6.0	0.0	NA
<b>Hemocytometer</b>		49	+/-	12.3	24	+/-	6.0	0.0	NA
<b>Makler</b>		50	+/-	12.5	25	+/-	6.3	0.0	NA
<b>Fixed Coverslip</b>		49	+/-	12.3	25	+/-	6.3	0.0	NA