

MEDICAL ELECTRONIC SYSTEMS

WWW.MES-GLOBAL.COM

QwikCheck™ Beads Batch Release Form									
BATCH #	061016	Level 1			Level 2			NEGATIVE CONTROL	
EXP DATE:	6-Oct-17	LOT# 061016001			LOT# 061016002			LOT# 061016003	
For use with:		Conc.			Conc.			Conc	MSC
SQA / QwikCheck gold Automated		48	+/-	6.7	25	+/-	5.0	0.0	0.0
SQA-V Visualization		48	+/-	12.0	25	+/-	6.3	0.0	NA
Hemocytometer		48	+/-	12.0	25	+/-	6.3	0.0	NA
Makler		49	+/-	12.3	27	+/-	6.8	0.0	NA
Fixed Coverslip		48	+/-	12.0	26	+/-	6.5	0.0	NA