



**QwikCheck-beads Batch Release Form**  
**QwikCheck-beads: Batch # 010411; Expiration Date: 04/2012**

Method	Level 1 Concentration M/ml	Level 2 Concentration M/ml	Negative Control	
			Concentration	MSC
	LOT# 010411001	LOT# 010411002	LOT# 010411003	
<b>SQA-V Automated</b>	45+/-6.3 Low 38.7 High 51.3 Range (±14%)	23+/-4.6 Low 18.4 High 27.6 Range (±20%)	0.0	0.0
<b>SQA-V Manual</b>	44+/-11.0 Low 33.0 High 55.0 Range (±25%)	23+/-5.8 Low 17.2 High 28.8 Range (±25%)	0.0	0.0
<b>Hemocytometer</b>	45+/-11.3 Low 33.7 High 56.3 Range (±25%)	23+/-5.8 Low 17.2 High 28.8 Range (±25%)	0.0	N.A.
<b>Makler</b>	45+/-11.3 Low 33.7 High 56.3 Range (±25%)	24+/-6.0 Low 18.0 High 30.0 Range (±25%)	0.0	N.A.
<b>Fixed Coverslip</b>	48+/-12.0 Low 36.0 High 60.0 Range (±25%)	23+/-5.8 Low 17.2 High 28.8 Range (±25%)	0.0	N.A.