



QwikCheck-beads Batch Release Form
QwikCheck-beads: Batch # 010611; Expiration Date: 06/2012

Method	Level 1 Concentration M/ml	Level 2 Concentration M/ml	Negative Control	
			Concentration	MSC
	LOT# 010611001	LOT# 010611002	LOT# 010611003	
SQA-V Automated	45+/-6.3 Low 38.7 High 51.3 Range (±14%)	23+/-4.6 Low 18.4 High 27.6 Range (±20%)	0.0	0.0
SQA-V Manual	43+/-10.8 Low 32.2 High 53.8 Range (±25%)	21+/-5.3 Low 15.7 High 26.3 Range (±25%)	0.0	0.0
Hemocytometer	46+/-11.5 Low 34.5 High 57.5 Range (±25%)	22+/-5.5 Low 16.5 High 27.5 Range (±25%)	0.0	N.A.
Makler	45+/-11.3 Low 33.7 High 56.3 Range (±25%)	22+/-5.5 Low 16.5 High 27.5 Range (±25%)	0.0	N.A.
Fixed Coverslip	45+/-11.3 Low 33.7 High 56.3 Range (±25%)	22+/-5.5 Low 16.5 High 27.5 Range (±25%)	0.0	N.A.