

MEDICAL ELECTRONIC SYSTEMS

WWW.MES-GLOBAL.COM

QwikCheck™ Beads Batch Release Form										
BATCH #	270817		Level 1			Level 2			NEGATIVE CONTROL	
EXP DATE:	27_August_2018		LOT#	270817001		LOT#	270817002		LOT#	270817003
For use with:	Conc.			Conc.			Conc	MSC		
SQA / QwikCheck gold Automated	49	+/-	6.9	26	+/-	5.2	0.0	0.0		
SQA-V Visualization	47	+/-	11.8	25	+/-	6.3	0.0	NA		
Hemocytometer	49	+/-	12.3	25	+/-	6.3	0.0	NA		
Makler	48	+/-	12.0	25	+/-	6.3	0.0	NA		
Fixed Coverslip	48	+/-	12.0	26	+/-	6.5	0.0	NA		