

MEDICAL ELECTRONIC SYSTEMS**WWW.MES-GLOBAL.COM**

QwikCheck™ Beads Batch Release Form									
BATCH #	180617	Level 1			Level 2			NEGATIVE CONTROL	
EXP DATE:	18_June_2018	LOT#	180617001	LOT#	180617002	LOT#	180617003		
For use with:	Conc.			Conc.			Conc	MSC	
SQA / QwikCheck gold Automated	47	+/-	6.6	25	+/-	5.0	0.0	0.0	
SQA-V Visualization	47	+/-	11.8	25	+/-	6.3	0.0	NA	
Hemocytometer	49	+/-	12.3	25	+/-	6.3	0.0	NA	
Makler	49	+/-	12.3	25	+/-	6.3	0.0	NA	
Fixed Coverslip	48	+/-	12.0	25	+/-	6.3	0.0	NA	